

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021524\
 Data File : VN081042.D
 Acq On : 15 Feb 2024 16:19
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 16 00:25:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021424W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 15 07:59:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	460820	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	850894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	783423	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	309677	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	285888	50.437	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.880%	
35) Dibromofluoromethane	8.165	113	239042	48.998	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.000%	
50) Toluene-d8	10.565	98	806340	46.865	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 93.720%	
62) 4-Bromofluorobenzene	12.853	95	322794	55.163	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.320%	
Target Compounds						Qvalue
13) Acrolein	4.171	56	1107	Below Cal	#	15
31) Cyclohexane	8.218	56	13062	1.327	ug/l #	7
36) 1,1-Dichloropropene	8.224	75	38944	4.784	ug/l #	53
38) Carbon Tetrachloride	8.224	117	49481	8.305	ug/l #	17
41) Methacrylonitrile	7.759	41	1223	0.271	ug/l #	47
44) Trichloroethene	9.365	130	1930	0.271	ug/l	90
49) 1,4-Dioxane	9.700	88	825	Below Cal	#	36
51) 4-Methyl-2-Pentanone	10.435	43	1886	0.242	ug/l #	79
55) 1,1,2-Trichloroethane	11.024	97	1280	0.226	ug/l #	19
67) Ethyl Benzene	11.941	91	6732	0.217	ug/l #	50
68) m/p-Xylenes	12.065	106	2855	0.250	ug/l #	58
70) Styrene	12.400	104	3685	0.209	ug/l #	41
76) 1,2,3-Trichloropropane	12.847	75	174327	25.477	ug/l #	1
81) trans-1,4-Dichloro-2-b...	12.847	75	174327	59.306	ug/l #	14
85) sec-Butylbenzene	13.606	105	7510	0.263	ug/l	80
86) p-Isopropyltoluene	13.753	119	10518	0.459	ug/l #	53
87) 1,3-Dichlorobenzene	13.723	146	5581	0.518	ug/l #	63
90) Hexachloroethane	14.394	117	851	Below Cal		84
92) 1,2-Dibromo-3-Chloropr...	14.988	75	632	Below Cal		80
93) 1,2,4-Trichlorobenzene	15.388	180	2602	0.416	ug/l #	18
95) Naphthalene	15.635	128	13298	0.597	ug/l #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021524\
Data File : VN081042.D
Acq On : 15 Feb 2024 16:19
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 16 00:25:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021424W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 15 07:59:19 2024
Response via : Initial Calibration

