

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021621\
 Data File : VN065874.D
 Acq On : 16 Feb 2021 15:26
 Operator : JC/MD
 Sample : VN0216WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0216WBL01

Quant Time: Feb 16 16:40:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 15 10:57:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	154953	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.544	114	259500	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	220583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	72220	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	84798	41.60	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	83.20%
35) Dibromofluoromethane	7.544	113	72962	47.45	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.90%
50) Toluene-d8	10.055	98	320323	52.99	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.98%
62) 4-Bromofluorobenzene	12.373	95	95868	44.04	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.08%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021621\
Data File : VN065874.D
Acq On : 16 Feb 2021 15:26
Operator : JC/MD
Sample : VN0216WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0216WBL01

Quant Time: Feb 16 16:40:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 15 10:57:01 2021
Response via : Initial Calibration

