

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021621\
 Data File : VN065876.D
 Acq On : 16 Feb 2021 16:34
 Operator : JC/MD
 Sample : M1413-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-021221

Quant Time: Feb 16 16:57:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 15 10:57:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	162217	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	272773	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	235497	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	83145	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	88553	41.49	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	82.98%
35) Dibromofluoromethane	7.544	113	77976	48.24	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.48%
50) Toluene-d8	10.055	98	335963	52.88	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.76%
62) 4-Bromofluorobenzene	12.373	95	106308	46.46	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.92%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021621\
Data File : VN065876.D
Acq On : 16 Feb 2021 16:34
Operator : JC/MD
Sample : M1413-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-021221

Quant Time: Feb 16 16:57:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 15 10:57:01 2021
Response via : Initial Calibration

