

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021621\
 Data File : VN065878.D
 Acq On : 16 Feb 2021 17:23
 Operator : JC/MD
 Sample : M1413-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-GW-021221

Quant Time: Feb 16 17:44:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 15 10:57:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	167758	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	282044	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	246767	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	87193	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	92609	41.96	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	83.92%
35) Dibromofluoromethane	7.547	113	80933	48.42	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.84%
50) Toluene-d8	10.058	98	347871	52.95	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.90%
62) 4-Bromofluorobenzene	12.373	95	110317	46.63	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.26%
Target Compounds						
16) Acetone	3.782	43	6338	7.38	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021621\
Data File : VN065878.D
Acq On : 16 Feb 2021 17:23
Operator : JC/MD
Sample : M1413-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-GW-021221

Quant Time: Feb 16 17:44:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 15 10:57:01 2021
Response via : Initial Calibration

