

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021621\
 Data File : VN065883.D
 Acq On : 16 Feb 2021 19:21
 Operator : JC/MD
 Sample : M1413-01DL 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWBR2-SC-93-021121DL

Quant Time: Feb 16 19:38:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 15 10:57:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	144991	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	242301	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	219419	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	83821	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	78806	41.31	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	82.62%
35) Dibromofluoromethane	7.547	113	69406	48.34	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.68%
50) Toluene-d8	10.058	98	301076	53.34	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.68%
62) 4-Bromofluorobenzene	12.373	95	101077	49.73	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.46%
Target Compounds						
						Qvalue
4) Vinyl Chloride	2.164	62	14394	5.96	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.714	101	2011	1.31	ug/l	94
16) Acetone	3.779	43	1863	2.51	ug/l	94
24) 1,1-Dichloroethane	5.795	63	8746	2.45	ug/l #	82
27) cis-1,2-Dichloroethene	6.775	96	144204	67.09	ug/l	91
30) Chloroform	7.325	83	1355	0.41	ug/l	95
32) 1,1,1-Trichloroethane	7.518	97	3217	1.20	ug/l #	46
52) Toluene	10.122	92	3083	0.70	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021621\
Data File : VN065883.D
Acq On : 16 Feb 2021 19:21
Operator : JC/MD
Sample : M1413-01DL 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWBR2-SC-93-021121DL

Quant Time: Feb 16 19:38:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 15 10:57:01 2021
Response via : Initial Calibration

