

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021622\
 Data File : VN070955.D
 Acq On : 16 Feb 2022 13:22
 Operator : JC\MD
 Sample : VN0216MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0216MBL01

Quant Time: Feb 17 04:08:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	982640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1489857	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1312883	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	527749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	533219	47.503	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.000%
35) Dibromofluoromethane	8.016	113	440680	49.729	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.460%
50) Toluene-d8	10.439	98	1748892	49.143	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.280%
62) 4-Bromofluorobenzene	12.727	95	567848	44.640	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.280%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021622\
Data File : VN070955.D
Acq On : 16 Feb 2022 13:22
Operator : JC\MD
Sample : VN0216MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0216MBL01

Quant Time: Feb 17 04:08:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

