

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021622\
 Data File : VN070959.D
 Acq On : 16 Feb 2022 15:24
 Operator : JC\MD
 Sample : N1573-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-12

Quant Time: Feb 17 05:13:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	867994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1331281	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1171888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	456143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	486677	49.084	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.160%
35) Dibromofluoromethane	8.016	113	390036	49.257	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.520%
50) Toluene-d8	10.439	98	1567897	49.305	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.620%
62) 4-Bromofluorobenzene	12.727	95	497130	43.736	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.480%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	66486	12.176	ug/l	97
37) Ethyl Acetate	7.416	43	50966	3.663	ug/l	100
89) n-Butylbenzene	13.939	91	14072	0.565	ug/l	99
93) 1,2,4-Trichlorobenzene	15.263	180	11549	3.319	ug/l	93
95) Naphthalene	15.498	128	43144	6.894	ug/l	99
96) 1,2,3-Trichlorobenzene	15.698	180	12734	4.078	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021622\
Data File : VN070959.D
Acq On : 16 Feb 2022 15:24
Operator : JC\MD
Sample : N1573-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-12

Quant Time: Feb 17 05:13:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

