

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021622\
 Data File : VN070961.D
 Acq On : 16 Feb 2022 16:13
 Operator : JC\MD
 Sample : N1573-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-20R

Quant Time: Feb 17 04:11:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	993509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1512945	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1342055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	514060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	546470	48.151	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.300%
35) Dibromofluoromethane	8.016	113	442318	49.152	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.300%
50) Toluene-d8	10.439	98	1776652	49.161	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.320%
62) 4-Bromofluorobenzene	12.727	95	571100	44.211	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.420%
Target Compounds						
					Qvalue	
16) Acetone	4.281	43	75081	12.013	ug/l	97
37) Ethyl Acetate	7.416	43	170483	10.782	ug/l	97
95) Naphthalene	15.504	128	10179	5.388	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021622\
Data File : VN070961.D
Acq On : 16 Feb 2022 16:13
Operator : JC\MD
Sample : N1573-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-20R

Quant Time: Feb 17 04:11:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

