

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021622\
 Data File : VN070963.D
 Acq On : 16 Feb 2022 17:02
 Operator : JC\MD
 Sample : N1573-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GES-2R

Quant Time: Feb 17 04:12:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	963147	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1476110	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1312817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	501834	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	542085	49.271	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.540%
35) Dibromofluoromethane	8.016	113	437856	49.870	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.740%
50) Toluene-d8	10.439	98	1752724	49.709	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.420%
62) 4-Bromofluorobenzene	12.727	95	557303	44.219	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.440%
Target Compounds						
					Qvalue	
16) Acetone	4.281	43	102626	16.937	ug/l	95
25) 2-Butanone	7.345	43	9652	1.197	ug/l #	88
37) Ethyl Acetate	7.410	43	43556	2.823	ug/l	99
95) Naphthalene	15.509	128	5121	5.195	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021622\
Data File : VN070963.D
Acq On : 16 Feb 2022 17:02
Operator : JC\MD
Sample : N1573-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GES-2R

Quant Time: Feb 17 04:12:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

