

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021622\
 Data File : VN070966.D
 Acq On : 16 Feb 2022 18:16
 Operator : JC\MD
 Sample : N1573-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-16SR

Quant Time: Feb 17 05:41:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	1018082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1562772	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1395758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	599308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	564254	48.518	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.040%
35) Dibromofluoromethane	8.016	113	454979	48.947	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.900%
50) Toluene-d8	10.439	98	1856252	49.726	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.460%
62) 4-Bromofluorobenzene	12.727	95	643891	48.256	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.520%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	82727	12.917	ug/l	91
22) Diisopropyl ether	6.487	45	20751	0.569	ug/l #	84
31) Cyclohexane	8.092	56	169271	8.416	ug/l #	92
37) Ethyl Acetate	7.416	43	124930	7.649	ug/l	98
39) Methylcyclohexane	9.457	83	51946	2.651	ug/l	94
40) Benzene	8.457	78	2931773	65.082	ug/l	100
52) Toluene	10.498	92	50252	1.775	ug/l	99
67) Ethyl Benzene	11.839	91	747220	14.195	ug/l	98
68) m/p-Xylenes	11.951	106	44983	2.212	ug/l	95
69) o-Xylene	12.274	106	39675	1.997	ug/l	91
73) Isopropylbenzene	12.574	105	973709	20.185	ug/l	100
78) n-propylbenzene	12.916	91	1420720	26.213	ug/l	98
84) 1,2,4-Trimethylbenzene	13.363	105	118846	3.089	ug/l	99
85) sec-Butylbenzene	13.498	105	56319	1.153	ug/l #	71
86) p-Isopropyltoluene	13.610	119	10745	0.268	ug/l #	77
95) Naphthalene	15.498	128	2504331	89.130	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021622\
Data File : VN070966.D
Acq On : 16 Feb 2022 18:16
Operator : JC\MD
Sample : N1573-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-16SR

Quant Time: Feb 17 05:41:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

