

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
 Data File : VN076701.D
 Acq On : 16 Feb 2023 10:53
 Operator : JC\MD
 Sample : VN0216MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0216MBL01

Quant Time: Feb 17 00:05:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	249062	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	462068	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	409341	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	162140	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.578	65	184722	53.711	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.420%
35) Dibromofluoromethane	8.172	113	145814	48.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	10.566	98	548457	48.533	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.060%
62) 4-Bromofluorobenzene	12.854	95	163084	43.572	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
Data File : VN076701.D
Acq On : 16 Feb 2023 10:53
Operator : JC\MD
Sample : VN0216MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0216MBL01

Quant Time: Feb 17 00:05:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

