

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
 Data File : VN076706.D
 Acq On : 16 Feb 2023 13:03
 Operator : JC\MD
 Sample : 01578-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230215017-01-TRIP-BLANK

Quant Time: Feb 17 00:09:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.237	168	284827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	517115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	446331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	165225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	206581	52.525	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.040%
35) Dibromofluoromethane	8.172	113	159061	46.969	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.940%
50) Toluene-d8	10.572	98	623964	49.337	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.680%
62) 4-Bromofluorobenzene	12.854	95	179160	42.772	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.540%
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	19228	10.945	ug/l	90
30) Chloroform	7.978	83	4412	0.674	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
Data File : VN076706.D
Acq On : 16 Feb 2023 13:03
Operator : JC\MD
Sample : 01578-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230215017-01-TRIP-BLANK

Quant Time: Feb 17 00:09:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

