

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
 Data File : VN076716.D
 Acq On : 16 Feb 2023 17:04
 Operator : JC\MD
 Sample : 01538-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TCLP-5

Quant Time: Feb 17 00:15:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	318095	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	590684	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	567141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	216705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	224856	51.192	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.380%
35) Dibromofluoromethane	8.178	113	173293	44.798	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.600%
50) Toluene-d8	10.572	98	747181	51.721	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.440%
62) 4-Bromofluorobenzene	12.854	95	234525	49.016	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.040%
Target Compounds						
						Qvalue
16) Acetone	4.454	43	38547	19.646	ug/l	97
20) Methylene Chloride	5.284	84	23694	5.762	ug/l #	82
25) 2-Butanone	7.501	43	16379	5.659	ug/l #	88
37) Ethyl Acetate	7.572	43	52853	8.988	ug/l	98
43) Isopropyl Acetate	8.695	43	554476	57.947	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
Data File : VN076716.D
Acq On : 16 Feb 2023 17:04
Operator : JC\MD
Sample : 01538-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TCPLP-5

Quant Time: Feb 17 00:15:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

