

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
 Data File : VN076717.D
 Acq On : 16 Feb 2023 17:28
 Operator : JC\MD
 Sample : 01538-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TCLP-6

Quant Time: Feb 17 00:15:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	299331	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	559308	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	545455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	206876	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	214620	51.924	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.840%
35) Dibromofluoromethane	8.178	113	168622	46.036	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.080%
50) Toluene-d8	10.572	98	711838	52.039	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.080%
62) 4-Bromofluorobenzene	12.854	95	220367	48.641	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.280%
Target Compounds						
						Qvalue
16) Acetone	4.454	43	35843	19.413	ug/l	90
20) Methylene Chloride	5.290	84	21932	5.668	ug/l	86
25) 2-Butanone	7.501	43	18202	6.683	ug/l	95
37) Ethyl Acetate	7.578	43	52877	9.496	ug/l	98
43) Isopropyl Acetate	8.695	43	369762	40.810	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
Data File : VN076717.D
Acq On : 16 Feb 2023 17:28
Operator : JC\MD
Sample : 01538-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TCPLP-6

Quant Time: Feb 17 00:15:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

