

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
 Data File : VN076719.D
 Acq On : 16 Feb 2023 18:16
 Operator : JC\MD
 Sample : 01539-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TCLP-8

Quant Time: Feb 17 00:16:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	303488	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	570029	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	545761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	207813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	216848	51.745	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.480%	
35) Dibromofluoromethane	8.172	113	165589	44.358	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 88.720%	
50) Toluene-d8	10.572	98	720447	51.678	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.360%	
62) 4-Bromofluorobenzene	12.854	95	221215	47.909	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.820%	
Target Compounds						
					Qvalue	
16) Acetone	4.449	43	39188	20.934	ug/l	97
20) Methylene Chloride	5.290	84	20627	5.258	ug/l	89
25) 2-Butanone	7.501	43	18431	6.675	ug/l	96
37) Ethyl Acetate	7.572	43	56020	9.872	ug/l	99
43) Isopropyl Acetate	8.695	43	498154	53.947	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
Data File : VN076719.D
Acq On : 16 Feb 2023 18:16
Operator : JC\MD
Sample : 01539-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TCPLP-8

Quant Time: Feb 17 00:16:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

