

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
 Data File : VN076720.D
 Acq On : 16 Feb 2023 18:40
 Operator : JC\MD
 Sample : PB150883TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150883TB

Quant Time: Feb 17 00:17:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	295510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	557442	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	549387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	206847	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	213608	52.348	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.700%			
35) Dibromofluoromethane	8.178	113	163894	44.895	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 89.800%			
50) Toluene-d8	10.572	98	705766	51.768	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.540%			
62) 4-Bromofluorobenzene	12.854	95	221758	49.112	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.220%			
Target Compounds						
					Qvalue	
16) Acetone	4.443	43	36199	19.860	ug/l	98
18) Methyl Acetate	5.048	43	6585	1.107	ug/l	93
20) Methylene Chloride	5.301	84	15800	4.136	ug/l #	81
37) Ethyl Acetate	7.566	43	60600	10.920	ug/l	98
43) Isopropyl Acetate	8.695	43	623391	69.034	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
Data File : VN076720.D
Acq On : 16 Feb 2023 18:40
Operator : JC\MD
Sample : PB150883TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150883TB

Quant Time: Feb 17 00:17:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

