

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
 Data File : VN076722.D
 Acq On : 16 Feb 2023 19:28
 Operator : JC\MD
 Sample : 01572-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RBR-200028

Quant Time: Feb 17 00:18:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	281597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	543406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	506388	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	195346	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	208444	53.606	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.220%		
35) Dibromofluoromethane	8.172	113	157574	44.279	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 88.560%		
50) Toluene-d8	10.571	98	675606	50.836	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.680%		
62) 4-Bromofluorobenzene	12.854	95	204041	46.355	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 92.720%		
Target Compounds						
						Qvalue
16) Acetone	4.454	43	43665	25.139	ug/l	97
18) Methyl Acetate	5.037	43	7627	1.345	ug/l	99
20) Methylene Chloride	5.284	84	21124	5.803	ug/l #	83
37) Ethyl Acetate	7.572	43	39460	7.294	ug/l #	95
43) Isopropyl Acetate	8.695	43	527290	59.900	ug/l #	71
95) Naphthalene	15.648	128	19080	1.821	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
Data File : VN076722.D
Acq On : 16 Feb 2023 19:28
Operator : JC\MD
Sample : 01572-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RBR-200028

Quant Time: Feb 17 00:18:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

