

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
 Data File : VN076723.D
 Acq On : 16 Feb 2023 19:52
 Operator : JC\MD
 Sample : 01549-02
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1695

Quant Time: Feb 17 00:18:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.225	168	308177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	568606	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	492269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	185898	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	217329	51.071	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.140%
35) Dibromofluoromethane	8.172	113	163666	43.953	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.900%
50) Toluene-d8	10.572	98	668177	48.048	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.100%
62) 4-Bromofluorobenzene	12.854	95	199620	43.341	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.680%
Target Compounds						
						Qvalue
18) Methyl Acetate	5.042	43	19924	3.211	ug/l	96
52) Toluene	10.630	92	51134	5.126	ug/l	96
68) m/p-Xylenes	12.071	106	11398	1.664	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
Data File : VN076723.D
Acq On : 16 Feb 2023 19:52
Operator : JC\MD
Sample : 01549-02
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1695

Quant Time: Feb 17 00:18:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

