

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
 Data File : VN076726.D
 Acq On : 16 Feb 2023 21:05
 Operator : JC\MD
 Sample : 01549-01
 Misc : 1.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1915

Quant Time: Feb 17 00:20:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.225	168	304682	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	543141	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	485374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	197890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	214477	50.978	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.960%	
35) Dibromofluoromethane	8.172	113	161967	45.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 91.080%	
50) Toluene-d8	10.571	98	666989	50.212	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.420%	
62) 4-Bromofluorobenzene	12.854	95	202969	46.134	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.260%	
Target Compounds						
					Qvalue	
16) Acetone	4.454	43	16830	8.955	ug/l	90
18) Methyl Acetate	5.060	43	16562	2.700	ug/l #	71
52) Toluene	10.636	92	48104	5.049	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021623\
Data File : VN076726.D
Acq On : 16 Feb 2023 21:05
Operator : JC\MD
Sample : 01549-01
Misc : 1.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1915

Quant Time: Feb 17 00:20:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

