

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021720\
 Data File : VN060116.D
 Acq On : 17 Feb 2020 13:59
 Operator : JC/MD
 Sample : PB126834ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126834ZHE#01

Quant Time: Feb 17 14:36:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	190001	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	305071	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	287469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	129160	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	120026	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
35) Dibromofluoromethane	7.57	113	92498	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
50) Toluene-d8	10.08	98	374460	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
62) 4-Bromofluorobenzene	12.40	95	137996	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
Target Compounds						
16) Acetone	3.79	43	5714	5.666	ug/l	96
20) Methylene Chloride	4.53	84	7234	3.595	ug/l	92
43) Isopropyl Acetate	8.14	43	58493	13.399	ug/l #	89
52) Toluene	10.14	92	1141	0.212	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021720\
Data File : VN060116.D
Acq On : 17 Feb 2020 13:59
Operator : JC/MD
Sample : PB126834ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126834ZHE#01

Quant Time: Feb 17 14:36:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

