

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021720\
 Data File : VN060119.D
 Acq On : 17 Feb 2020 15:07
 Operator : JC/MD
 Sample : PB126834ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126834ZHE#04

Quant Time: Feb 17 17:08:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	180682	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	290748	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	269928	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	119817	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	111935	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
35) Dibromofluoromethane	7.57	113	88082	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
50) Toluene-d8	10.08	98	354531	50.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.74%	
62) 4-Bromofluorobenzene	12.40	95	127657	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
Target Compounds						
16) Acetone	3.79	43	5439	5.671	ug/l	100
20) Methylene Chloride	4.53	84	7510	3.925	ug/l	95
43) Isopropyl Acetate	8.14	43	55809	13.414	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021720\
Data File : VN060119.D
Acq On : 17 Feb 2020 15:07
Operator : JC/MD
Sample : PB126834ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126834ZHE#04

Quant Time: Feb 17 17:08:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

