

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021720\
 Data File : VN060125.D
 Acq On : 17 Feb 2020 17:23
 Operator : JC/MD
 Sample : L1530-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IDW-GS-200213

Quant Time: Feb 18 02:44:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	178722	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	286858	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	272674	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	124972	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	110560	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
35) Dibromofluoromethane	7.57	113	87825	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
50) Toluene-d8	10.08	98	351760	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
62) 4-Bromofluorobenzene	12.40	95	130513	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
Target Compounds						
16) Acetone	3.79	43	11728	12.363	ug/l	98
18) Methyl Acetate	4.29	43	3095	1.737	ug/l #	80
20) Methylene Chloride	4.53	84	10804	5.709	ug/l	96
43) Isopropyl Acetate	8.14	43	101850	24.812	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021720\
Data File : VN060125.D
Acq On : 17 Feb 2020 17:23
Operator : JC/MD
Sample : L1530-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IDW-GS-200213

Quant Time: Feb 18 02:44:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

