

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021721\
 Data File : VN065925.D
 Acq On : 18 Feb 2021 4:37
 Operator : JC/MD
 Sample : M1418-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT-MW178I-GW-20210210

Quant Time: Feb 18 05:00:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	178240	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	303927	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	265750	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	86945	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	117333	47.07	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.14%
35) Dibromofluoromethane	7.553	113	89909	47.80	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.60%
50) Toluene-d8	10.058	98	374781	48.54	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.08%
62) 4-Bromofluorobenzene	12.373	95	114707	39.57	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	79.14%
Target Compounds						
19) Methyl tert-butyl Ether	5.000	73	2974	0.46	ug/l #	88
44) Trichloroethene	8.804	130	1425	0.63	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021721\
Data File : VN065925.D
Acq On : 18 Feb 2021 4:37
Operator : JC/MD
Sample : M1418-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT-MW178I-GW-20210210

Quant Time: Feb 18 05:00:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

