

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021721\
 Data File : VN065913.D
 Acq On : 17 Feb 2021 23:41
 Operator : JC/MD
 Sample : M1405-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Feb 18 01:49:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	196261	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	333825	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	300273	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	106268	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	127931	46.61	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.22%
35) Dibromofluoromethane	7.553	113	98701	47.77	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.54%
50) Toluene-d8	10.058	98	414043	48.82	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.64%
62) 4-Bromofluorobenzene	12.373	95	137580	43.21	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.42%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021721\
Data File : VN065913.D
Acq On : 17 Feb 2021 23:41
Operator : JC/MD
Sample : M1405-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Feb 18 01:49:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.627 min Scan# 1831
Delta R.T. 0.000 min
Lab File: VN065913.D
Acq: 17 Feb 2021 23:41




Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:168 Resp: 196261
Ion Ratio Lower Upper
168 100
99 61.7 42.6 63.8



#33
1,2-Dichloroethane-d4
Concen: 46.61 ug/l
RT: 7.987 min Scan# 1943
Delta R.T. 0.000 min
Lab File: VN065913.D
Acq: 17 Feb 2021 23:41

Tgt Ion: 65 Resp: 127931
Ion Ratio Lower Upper
65 100
67 53.1 0.0 107.2









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.550 min Scan# 2118
Delta R.T. 0.000 min
Lab File: VN065913.D
Acq: 17 Feb 2021 23:41




Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:114 Resp: 333825
Ion Ratio Lower Upper
114 100
63 21.5 0.0 38.6
88 16.5 0.0 30.6



#35
Dibromofluoromethane
Concen: 47.77 ug/l
RT: 7.553 min Scan# 1808
Delta R.T. 0.003 min
Lab File: VN065913.D
Acq: 17 Feb 2021 23:41

Tgt Ion:113 Resp: 98701
Ion Ratio Lower Upper
113 100
111 104.2 82.2 123.2
192 17.9 15.4 23.0









#50
Toluene-d8
Concen: 48.82 ug/l
RT: 10.058 min Scan# 2587
Delta R.T. 0.000 min
Lab File: VN065913.D
Acq: 17 Feb 2021 23:41




Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion: 98 Resp: 414043
Ion Ratio Lower Upper
98 100
100 64.6 50.5 75.7



#62
4-Bromofluorobenzene
Concen: 43.21 ug/l
RT: 12.373 min Scan# 3307
Delta R.T. 0.003 min
Lab File: VN065913.D
Acq: 17 Feb 2021 23:41

Tgt Ion: 95 Resp: 137580
Ion Ratio Lower Upper
95 100
174 69.9 0.0 169.0
176 67.6 0.0 162.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.376 min Scan# 2997
Delta R.T. 0.000 min
Lab File: VN065913.D
Acq: 17 Feb 2021 23:41

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Tgt Ion:	117	Resp:	300273
Ion Ratio	Lower	Upper	
117	100		
82	57.7	43.7	65.5
119	31.9	26.1	39.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.315 min Scan# 3600
Delta R.T. 0.003 min
Lab File: VN065913.D
Acq: 17 Feb 2021 23:41

Tgt Ion:	152	Resp:	106268
Ion Ratio	Lower	Upper	
152	100		
115	61.9	28.5	85.6
150	157.7	0.0	347.6

