

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
 Data File : VN070974.D
 Acq On : 17 Feb 2022 12:50
 Operator : JC\MD
 Sample : VN0217WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0217WBL01

Quant Time: Feb 18 00:43:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	887290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1380293	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1210880	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	446397	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	509275	50.246	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.500%	
35) Dibromofluoromethane	8.016	113	411133	50.077	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.160%	
50) Toluene-d8	10.439	98	1637122	49.654	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.300%	
62) 4-Bromofluorobenzene	12.727	95	550189	46.685	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 93.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
Data File : VN070974.D
Acq On : 17 Feb 2022 12:50
Operator : JC\MD
Sample : VN0217WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0217WBL01

Quant Time: Feb 18 00:43:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

