

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
 Data File : VN070976.D
 Acq On : 17 Feb 2022 13:49
 Operator : JC\MD
 Sample : N1581-24 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP01-20220215

Quant Time: Feb 18 00:44:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	819654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1284383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1127323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	399057	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	483071	51.593	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.180%
35) Dibromofluoromethane	8.022	113	382052	50.010	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.020%
50) Toluene-d8	10.439	98	1521916	49.607	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.220%
62) 4-Bromofluorobenzene	12.727	95	448755	40.921	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	81.840%
Target Compounds						
					Qvalue	
59) 2-Hexanone	11.092	43	10233	1.156	ug/l	99
93) 1,2,4-Trichlorobenzene	15.268	180	6705	2.875	ug/l	99
95) Naphthalene	15.504	128	27485	6.376	ug/l	99
96) 1,2,3-Trichlorobenzene	15.704	180	9377	3.831	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
Data File : VN070976.D
Acq On : 17 Feb 2022 13:49
Operator : JC\MD
Sample : N1581-24 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP01-20220215

Quant Time: Feb 18 00:44:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

