

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
 Data File : VN070977.D
 Acq On : 17 Feb 2022 14:14
 Operator : JC\MD
 Sample : N1581-25 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP02-20220215

Quant Time: Feb 18 00:44:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	831605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1289612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1139616	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	411177	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	478409	50.361	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.720%	
35) Dibromofluoromethane	8.022	113	386455	50.381	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.760%	
50) Toluene-d8	10.439	98	1518227	49.286	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.580%	
62) 4-Bromofluorobenzene	12.727	95	457257	41.528	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 83.060%	
Target Compounds						
						Qvalue
93) 1,2,4-Trichlorobenzene	15.268	180	2547	2.322	ug/l	95
95) Naphthalene	15.504	128	10820	5.519	ug/l	97
96) 1,2,3-Trichlorobenzene	15.704	180	2982	2.922	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
Data File : VN070977.D
Acq On : 17 Feb 2022 14:14
Operator : JC\MD
Sample : N1581-25 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP02-20220215

Quant Time: Feb 18 00:44:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

