

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
 Data File : VN070978.D
 Acq On : 17 Feb 2022 14:39
 Operator : JC\MD
 Sample : N1581-20 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 INTERCEPTOR-SUMP-20220215

Quant Time: Feb 18 00:45:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	854772	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1335582	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1174592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	423157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	498842	51.089	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.180%	
35) Dibromofluoromethane	8.016	113	402307	50.643	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.280%	
50) Toluene-d8	10.439	98	1585493	49.698	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.400%	
62) 4-Bromofluorobenzene	12.727	95	477394	41.864	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 83.720%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.322	96	166985	16.176	ug/l	96
44) Trichloroethene	9.216	130	792882	78.211	ug/l	100
95) Naphthalene	15.509	128	6118	5.281	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
Data File : VN070978.D
Acq On : 17 Feb 2022 14:39
Operator : JC\MD
Sample : N1581-20 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
INTERCEPTOR-SUMP-20220215

Quant Time: Feb 18 00:45:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

