

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
 Data File : VN070985.D
 Acq On : 17 Feb 2022 17:33
 Operator : JC\MD
 Sample : N1581-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-118.5T-20220215

Quant Time: Feb 18 00:48:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	856617	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1342566	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1171730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	397342	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	500855	51.185	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.360%
35) Dibromofluoromethane	8.022	113	398626	49.918	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.840%
50) Toluene-d8	10.439	98	1600982	49.922	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.840%
62) 4-Bromofluorobenzene	12.727	95	503149	43.893	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.780%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	20387	3.783	ug/l	# 87
27) cis-1,2-Dichloroethene	7.322	96	29785	2.879	ug/l	93
44) Trichloroethene	9.210	130	23079	2.265	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
Data File : VN070985.D
Acq On : 17 Feb 2022 17:33
Operator : JC\MD
Sample : N1581-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-118.5T-20220215

Quant Time: Feb 18 00:48:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

