

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
 Data File : VN070986.D
 Acq On : 17 Feb 2022 17:58
 Operator : JC\MD
 Sample : N1581-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-121.5N-20220215

Quant Time: Feb 18 00:48:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	917369	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1425506	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1274088	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	426925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	530771	50.650	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.300%
35) Dibromofluoromethane	8.022	113	421595	49.723	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.440%
50) Toluene-d8	10.439	98	1709615	50.208	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.420%
62) 4-Bromofluorobenzene	12.727	95	554059	45.522	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.040%
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.322	96	4164	0.376	ug/l	87
44) Trichloroethene	9.210	130	3156	0.292	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
Data File : VN070986.D
Acq On : 17 Feb 2022 17:58
Operator : JC\MD
Sample : N1581-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-121.5N-20220215

Quant Time: Feb 18 00:48:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

