

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
 Data File : VN070987.D
 Acq On : 17 Feb 2022 18:23
 Operator : JC\MD
 Sample : N1581-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-121.5T-20220215

Quant Time: Feb 18 00:49:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	839163	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1300496	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1149866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	385377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	489095	51.022	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.040%	
35) Dibromofluoromethane	8.022	113	389824	50.395	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.800%	
50) Toluene-d8	10.439	98	1557980	50.153	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.300%	
62) 4-Bromofluorobenzene	12.727	95	495882	44.659	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 89.320%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	22206	4.206	ug/l	99
27) cis-1,2-Dichloroethene	7.322	96	70582	6.964	ug/l	96
44) Trichloroethene	9.210	130	52814	5.350	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
Data File : VN070987.D
Acq On : 17 Feb 2022 18:23
Operator : JC\MD
Sample : N1581-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-121.5T-20220215

Quant Time: Feb 18 00:49:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

