

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
 Data File : VN070990.D
 Acq On : 17 Feb 2022 19:38
 Operator : JC\MD
 Sample : N1581-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-388.9-20220215

Quant Time: Feb 18 00:50:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	831250	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1319912	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1158538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	402956	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	492429	51.859	ug/l	0.01
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.720%	
35) Dibromofluoromethane	8.022	113	389684	49.636	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.280%	
50) Toluene-d8	10.439	98	1586921	50.333	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.660%	
62) 4-Bromofluorobenzene	12.727	95	499535	44.326	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 88.660%	
Target Compounds						
16) Acetone	4.292	43	19201	3.672	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
Data File : VN070990.D
Acq On : 17 Feb 2022 19:38
Operator : JC\MD
Sample : N1581-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-388.9-20220215

Quant Time: Feb 18 00:50:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

