


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021722\
 Data File : VN070972.D
 Acq On : 17 Feb 2022 10:57
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Title : SW846 8260
 Last Update : Mon Feb 21 07:57:15 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	19341	PASS
75	95	30	60	50.6	50997	PASS
95	95	100	100	100.0	100763	PASS
96	95	5	9	7.2	7209	PASS
173	174	0.00	2	1.2	979	PASS
174	95	50	100	82.0	82640	PASS
175	174	5	9	7.3	6023	PASS
176	174	95	101	95.6	78997	PASS
177	176	5	9	6.6	5239	PASS