

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021723\
 Data File : VN076735.D
 Acq On : 17 Feb 2023 15:28
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 18 02:09:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N021523W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Feb 16 01:08:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.819	128	27142	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.107	114	185899	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	196662	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.589	65	81028	32.640	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	108.800%
60) 4-Bromofluorobenzene	12.854	95	83471	23.996	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	80.000%
63) Toluene-d8	10.572	98	204127	29.633	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.767%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021723\
Data File : VN076735.D
Acq On : 17 Feb 2023 15:28
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 18 02:09:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N021523W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Feb 16 01:08:46 2023
Response via : Initial Calibration

