

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021723\
 Data File : VN076736.D
 Acq On : 17 Feb 2023 15:52
 Operator : JC\MD
 Sample : 01547-02RE 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-2RE

Quant Time: Feb 18 02:09:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N021523W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Feb 16 01:08:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.813	128	11587	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.107	114	173731	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.866	117	176625	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.589	65	76870	72.535	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	241.800%#
60) 4-Bromofluorobenzene	12.854	95	79668	25.501	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	85.000%
63) Toluene-d8	10.572	98	187391	30.290	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.967%
Target Compounds						
25) Chloroform	7.972	83	50884	20.496	ug/l	96
41) Bromodichloromethane	9.895	83	7366	1.751	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021723\
Data File : VN076736.D
Acq On : 17 Feb 2023 15:52
Operator : JC\MD
Sample : 01547-02RE 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-2RE

Quant Time: Feb 18 02:09:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N021523W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Feb 16 01:08:46 2023
Response via : Initial Calibration

