

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021819\
 Data File : VN053844.D
 Acq On : 18 Feb 2019 15:41
 Operator : JC/SP
 Sample : PB117159ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117159ZHE#08

Quant Time: Feb 19 03:43:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	672404	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1074233	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1029214	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	407528	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	363526	53.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.78%	
35) Dibromofluoromethane	7.59	113	360644	53.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.26%	
50) Toluene-d8	10.09	98	1324669	51.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.94%	
62) 4-Bromofluorobenzene	12.40	95	432822	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
Target Compounds						
16) Acetone	3.82	43	24354	14.075	ug/l	95
20) Methylene Chloride	4.56	84	11693	1.611	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021819\
Data File : VN053844.D
Acq On : 18 Feb 2019 15:41
Operator : JC/SP
Sample : PB117159ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117159ZHE#08

Quant Time: Feb 19 03:43:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

