

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021819\
 Data File : VN053852.D
 Acq On : 18 Feb 2019 19:15
 Operator : JC/SP
 Sample : PB117159ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117159ZHE#16

Quant Time: Feb 19 03:56:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	613116	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1049431	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1008090	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	391780	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	334951	54.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.92%	
35) Dibromofluoromethane	7.59	113	335681	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
50) Toluene-d8	10.09	98	1294459	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
62) 4-Bromofluorobenzene	12.40	95	422080	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
Target Compounds						
16) Acetone	3.82	43	21056	13.346	ug/l	98
20) Methylene Chloride	4.55	84	9200	1.390	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021819\
Data File : VN053852.D
Acq On : 18 Feb 2019 19:15
Operator : JC/SP
Sample : PB117159ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117159ZHE#16

Quant Time: Feb 19 03:56:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

