

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN021819\
 Data File : VN053830.D
 Acq On : 18 Feb 2019 9:02
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 19 03:15:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.515	0.451	12.4	82	0.00
3 P	Chloromethane	0.594	0.487	18.0	84	0.00
4 C	Vinyl Chloride	0.576	0.512	11.1#	89	0.00
5 T	Bromomethane	0.377	0.342	9.3	94	0.00
6 T	Chloroethane	0.329	0.309	6.1	97	0.00
7 T	Trichlorofluoromethane	0.781	0.753	3.6	99	0.00
8 T	Diethyl Ether	0.285	0.264	7.4	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.482	0.2	104	0.00
10 T	Methyl Iodide	0.759	0.682	10.1	92	0.00
11 T	Tert butyl alcohol	0.030	0.031	-3.3	108	0.00
12 CM	1,1-Dichloroethene	0.463	0.430	7.1#	95	0.00
13 T	Acrolein	0.016	0.024	-50.0#	188#	0.00
14 T	Allyl chloride	0.737	0.684	7.2	94	0.00
15 T	Acrylonitrile	0.160	0.164	-2.5	103	0.00
16 T	Acetone	0.129	0.173	-34.1#	147	0.00
17 T	Carbon Disulfide	1.442	1.165	19.2	83	0.00
18 T	Methyl Acetate	0.381	0.420	-10.2	118	0.00
19 T	Methyl tert-butyl Ether	1.174	1.179	-0.4	100	0.00
20 T	Methylene Chloride	0.540	0.500	7.4	100	0.00
21 T	trans-1,2-Dichloroethene	0.500	0.459	8.2	95	0.00
22 T	Diisopropyl ether	1.366	1.369	-0.2	100	0.00
23 T	Vinyl Acetate	0.945	0.910	3.7	94	0.00
24 P	1,1-Dichloroethane	0.911	0.862	5.4	98	0.00
25 T	2-Butanone	0.183	0.199	-8.7	111	0.00
26 T	2,2-Dichloropropane	0.696	0.686	1.4	102	0.00
27 T	cis-1,2-Dichloroethene	0.555	0.537	3.2	98	0.00
28 T	Bromochloromethane	0.395	0.387	2.0	101	0.00
29 T	Tetrahydrofuran	0.118	0.115	2.5	97	0.00
30 C	Chloroform	0.903	0.871	3.5#	102	0.00
31 T	Cyclohexane	0.949	0.738	22.2	90	0.00
32 T	1,1,1-Trichloroethane	0.752	0.730	2.9	100	0.00
33 S	1,2-Dichloroethane-d4	0.502	0.494	1.6	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.316	0.328	-3.8	103	0.00
36 T	1,1-Dichloropropene	0.451	0.438	2.9	97	0.00
37 T	Ethyl Acetate	0.252	0.251	0.4	92	0.00
38 T	Carbon Tetrachloride	0.454	0.445	2.0	98	0.00
39 T	Methylcyclohexane	0.535	0.529	1.1	95	0.00
40 TM	Benzene	1.390	1.346	3.2	97	0.00
41 T	Methacrylonitrile	0.141	0.160	-13.5	114	0.02
42 TM	1,2-Dichloroethane	0.413	0.401	2.9	99	0.00
43 T	Isopropyl Acetate	0.486	0.451	7.2	95	0.00
44 TM	Trichloroethene	0.396	0.384	3.0	101	0.00
45 C	1,2-Dichloropropane	0.356	0.356	0.0	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN021819\
 Data File : VN053830.D
 Acq On : 18 Feb 2019 9:02
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 19 03:15:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.214	0.211	1.4	100	0.00
47 T	Bromodichloromethane	0.446	0.451	-1.1	102	0.00
48 T	Methyl methacrylate	0.232	0.219	5.6	95	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	107	0.00
50 S	Toluene-d8	1.186	1.201	-1.3	101	0.00
51 T	4-Methyl-2-Pentanone	0.239	0.250	-4.6	100	0.00
52 CM	Toluene	0.825	0.823	0.2#	98	0.00
53 T	t-1,3-Dichloropropene	0.446	0.455	-2.0	100	0.00
54 T	cis-1,3-Dichloropropene	0.529	0.535	-1.1	99	0.00
55 T	1,1,2-Trichloroethane	0.297	0.308	-3.7	102	0.00
56 T	Ethyl methacrylate	0.352	0.360	-2.3	97	0.00
57 T	1,3-Dichloropropane	0.508	0.503	1.0	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.196	0.165	15.8	78	0.00
59 T	2-Hexanone	0.162	0.170	-4.9	101	0.00
60 T	Dibromochloromethane	0.335	0.348	-3.9	102	0.00
61 T	1,2-Dibromoethane	0.294	0.292	0.7	100	0.00
62 S	4-Bromofluorobenzene	0.396	0.381	3.8	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.429	0.430	-0.2	100	0.00
65 PM	Chlorobenzene	1.062	1.038	2.3	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.386	-4.0	102	0.00
67 C	Ethyl Benzene	1.745	1.731	0.8#	97	0.00
68 T	m/p-Xylenes	0.666	0.679	-2.0	98	0.00
69 T	o-Xylene	0.643	0.657	-2.2	98	0.00
70 T	Styrene	1.043	1.077	-3.3	98	0.00
71 P	Bromoform	0.254	0.277	-9.1	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.515	3.664	-4.2	98	0.00
74 T	N-amyl acetate	0.845	0.880	-4.1	96	0.00
75 P	1,1,2,2-Tetrachloroethane	0.841	0.801	4.8	93	0.00
76 T	1,2,3-Trichloropropane	0.642	0.701	-9.2	105	0.00
77 T	Bromobenzene	0.947	0.929	1.9	93	0.00
78 T	n-propylbenzene	4.063	4.101	-0.9	94	0.00
79 T	2-Chlorotoluene	2.411	2.387	1.0	94	0.00
80 T	1,3,5-Trimethylbenzene	2.915	3.024	-3.7	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.228	0.230	-0.9	92	0.00
82 T	4-Chlorotoluene	2.474	2.410	2.6	92	0.00
83 T	tert-Butylbenzene	2.576	2.663	-3.4	97	0.00
84 T	1,2,4-Trimethylbenzene	2.970	3.080	-3.7	95	0.00
85 T	sec-Butylbenzene	3.506	3.657	-4.3	97	0.00
86 T	p-Isopropyltoluene	3.066	3.206	-4.6	96	0.00
87 T	1,3-Dichlorobenzene	1.703	1.692	0.6	96	0.00
88 T	1,4-Dichlorobenzene	1.685	1.647	2.3	94	0.00
89 T	n-Butylbenzene	2.654	2.768	-4.3	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN021819\
Data File : VN053830.D
Acq On : 18 Feb 2019 9:02
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 19 03:15:21 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.506	0.553	-9.3	105	0.00
91 T	1,2-Dichlorobenzene	1.630	1.621	0.6	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.121	-0.8	94	0.00
93 T	1,2,4-Trichlorobenzene	0.935	0.933	0.2	87	0.00
94 T	Hexachlorobutadiene	0.582	0.647	-11.2	104	0.00
95 T	Naphthalene	1.905	1.758	7.7	76	0.00
96 T	1,2,3-Trichlorobenzene	0.868	0.869	-0.1	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5