

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021821\
 Data File : VN065935.D
 Acq On : 18 Feb 2021 12:54
 Operator : JC/MD
 Sample : M1448-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWBR3-95-113-021621

Quant Time: Feb 19 00:15:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	175436	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	300294	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	270470	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	97803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	127901	52.13	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	104.26%	
35) Dibromofluoromethane	7.553	113	99212	53.38	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	106.76%	
50) Toluene-d8	10.058	98	402172	52.72	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	105.44%	
62) 4-Bromofluorobenzene	12.373	95	135401	47.28	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	94.56%	
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	3.701	96	40567	20.59	ug/l	97
16) Acetone	3.791	43	4282	3.98	ug/l	92
21) trans-1,2-Dichloroethene	4.984	96	2803	1.35	ug/l #	78
24) 1,1-Dichloroethane	5.804	63	6186	1.59	ug/l #	92
27) cis-1,2-Dichloroethene	6.785	96	24006	10.18	ug/l	96
30) Chloroform	7.335	83	26862	6.99	ug/l	94
42) 1,2-Dichloroethane	8.090	62	3980	1.33	ug/l	90
44) Trichloroethene	8.801	130	199364	88.73	ug/l	98
47) Bromodichloromethane	9.367	83	3900	1.35	ug/l #	88
52) Toluene	10.122	92	65693	11.77	ug/l	96
60) Dibromochloromethane	10.872	129	881	0.41	ug/l	84
64) Tetrachloroethene	10.601	164	904	0.43	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021821\
 Data File : VN065935.D
 Acq On : 18 Feb 2021 12:54
 Operator : JC/MD
 Sample : M1448-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
Client Sampled :
 GWBR3-95-113-021621

Quant Time: Feb 19 00:15:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

