

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021821\
 Data File : VN065937.D
 Acq On : 18 Feb 2021 13:46
 Operator : JC/MD
 Sample : M1448-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWBR3-138-150-021721

Quant Time: Feb 19 00:15:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	181265	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	304517	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	279131	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	96950	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	129431	51.06	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.12%
35) Dibromofluoromethane	7.550	113	98318	52.17	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.34%
50) Toluene-d8	10.058	98	405476	52.41	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.82%
62) 4-Bromofluorobenzene	12.373	95	135022	46.49	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.98%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	3.702	96	5694	2.80	ug/l	99
27) cis-1,2-Dichloroethene	6.791	96	4591	1.88	ug/l	90
30) Chloroform	7.338	83	1667	0.42	ug/l	# 77
44) Trichloroethene	8.804	130	10264	4.50	ug/l	99
52) Toluene	10.122	92	15584	2.75	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021821\
Data File : VN065937.D
Acq On : 18 Feb 2021 13:46
Operator : JC/MD
Sample : M1448-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWBR3-138-150-021721

Quant Time: Feb 19 00:15:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

