

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
 Data File : VN070998.D
 Acq On : 18 Feb 2022 11:39
 Operator : JC\MD
 Sample : VN0218MBL01
 Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0218MBL01

Quant Time: Feb 19 00:37:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	789893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1252816	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1107091	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	366275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	474858	52.627	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.260%	
35) Dibromofluoromethane	8.022	113	381317	51.171	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.340%	
50) Toluene-d8	10.439	98	1506592	50.345	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.680%	
62) 4-Bromofluorobenzene	12.727	95	434573	40.627	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 81.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
Data File : VN070998.D
Acq On : 18 Feb 2022 11:39
Operator : JC\MD
Sample : VN0218MBL01
Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0218MBL01

Quant Time: Feb 19 00:37:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

