

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
 Data File : VN071001.D
 Acq On : 18 Feb 2022 15:26
 Operator : JC\MD
 Sample : N1581-20 25X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 INTERCEPTOR-SUMP-20220215

Quant Time: Feb 19 00:40:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	835928	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1297863	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1144965	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	386958	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	489191	51.230	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.460%	
35) Dibromofluoromethane	8.016	113	388543	50.331	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.660%	
50) Toluene-d8	10.439	98	1554474	50.142	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.280%	
62) 4-Bromofluorobenzene	12.727	95	511690	46.176	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 92.360%	
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.322	96	265026	26.252	ug/l	97
44) Trichloroethene	9.216	130	1213100	123.140	ug/l	99
59) 2-Hexanone	11.086	43	14698	1.643	ug/l	96
93) 1,2,4-Trichlorobenzene	15.262	180	7881	3.061	ug/l	92
95) Naphthalene	15.504	128	27574	6.424	ug/l	98
96) 1,2,3-Trichlorobenzene	15.704	180	8898	3.803	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
Data File : VN071001.D
Acq On : 18 Feb 2022 15:26
Operator : JC\MD
Sample : N1581-20 25X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
INTERCEPTOR-SUMP-20220215

Quant Time: Feb 19 00:40:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

