

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
 Data File : VN071003.D
 Acq On : 18 Feb 2022 16:14
 Operator : JC\MD
 Sample : N1581-24
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP01-20220215

Quant Time: Feb 19 00:40:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	848691	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1335115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1165209	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	382586	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	503620	51.948	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.900%
35) Dibromofluoromethane	8.022	113	394206	49.640	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.280%
50) Toluene-d8	10.439	98	1609913	50.481	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.960%
62) 4-Bromofluorobenzene	12.727	95	511925	44.908	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.820%
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	13186	2.470	ug/l	100
44) Trichloroethene	9.210	130	5744	0.567	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
Data File : VN071003.D
Acq On : 18 Feb 2022 16:14
Operator : JC\MD
Sample : N1581-24
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP01-20220215

Quant Time: Feb 19 00:40:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

