

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
 Data File : VN071004.D
 Acq On : 18 Feb 2022 16:38
 Operator : JC\MD
 Sample : N1581-25
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP02-20220215

Quant Time: Mar 02 13:27:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	859448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1350505	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1179017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	384786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	502021	51.135	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.260%	
35) Dibromofluoromethane	8.022	113	396683	49.383	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.760%	
50) Toluene-d8	10.439	98	1620042	50.220	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.440%	
62) 4-Bromofluorobenzene	12.727	95	511257	44.338	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 88.680%	
Target Compounds						
						Qvalue
16) Acetone	4.304	43	17758	3.284	ug/l	97
43) Isopropyl Acetate	8.657	43	8644	0.389	ug/l #	55
44) Trichloroethene	9.216	130	15626	1.524	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
Data File : VN071004.D
Acq On : 18 Feb 2022 16:38
Operator : JC\MD
Sample : N1581-25
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP02-20220215

Quant Time: Mar 02 13:27:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

