

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
 Data File : VN071005.D
 Acq On : 18 Feb 2022 17:03
 Operator : JC\MD
 Sample : N1581-21
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ROOF-DRAIN-20220215

Quant Time: Feb 19 00:41:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	881445	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1380592	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1211019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	405792	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	513870	51.035	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.080%
35) Dibromofluoromethane	8.022	113	404102	49.210	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.420%
50) Toluene-d8	10.439	98	1667145	50.554	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.100%
62) 4-Bromofluorobenzene	12.727	95	529655	44.933	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.860%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	22259	4.014	ug/l	97
27) cis-1,2-Dichloroethene	7.328	96	7261	0.682	ug/l	94
44) Trichloroethene	9.216	130	18034	1.721	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
Data File : VN071005.D
Acq On : 18 Feb 2022 17:03
Operator : JC\MD
Sample : N1581-21
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ROOF-DRAIN-20220215

Quant Time: Feb 19 00:41:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

