

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
 Data File : VN071009.D
 Acq On : 18 Feb 2022 18:39
 Operator : JC\MD
 Sample : N1587-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Quant Time: Feb 19 02:54:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 16 07:52:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	862274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1352417	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1189776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	400657	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	512113	51.992	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.980%	
35) Dibromofluoromethane	8.016	113	401285	49.885	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.780%	
50) Toluene-d8	10.439	98	1644683	50.911	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.820%	
62) 4-Bromofluorobenzene	12.727	95	527046	45.643	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 91.280%	
Target Compounds						
					Qvalue	
3) Chloromethane	2.299	50	3468	0.308	ug/l #	77
16) Acetone	4.304	43	11501	2.120	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN021822\
Data File : VN071009.D
Acq On : 18 Feb 2022 18:39
Operator : JC\MD
Sample : N1587-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Quant Time: Feb 19 02:54:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 16 07:52:42 2022
Response via : Initial Calibration

