

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021919\
 Data File : VN053861.D
 Acq On : 19 Feb 2019 11:34
 Operator : JC/SP
 Sample : K1525-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M-14-S

Quant Time: Feb 20 05:48:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	623174	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1041127	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	987829	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	388225	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	343094	54.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.76%	
35) Dibromofluoromethane	7.59	113	331004	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
50) Toluene-d8	10.09	98	1270614	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	12.40	95	417398	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
Target Compounds						
16) Acetone	3.82	43	22129	13.800	ug/l	99
20) Methylene Chloride	4.55	84	126473	18.799	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021919\
Data File : VN053861.D
Acq On : 19 Feb 2019 11:34
Operator : JC/SP
Sample : K1525-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M-14-S

Quant Time: Feb 20 05:48:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

