

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021919\
 Data File : VN053862.D
 Acq On : 19 Feb 2019 12:01
 Operator : JC/SP
 Sample : K1525-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 M-14-D

Quant Time: Feb 20 05:48:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	677692	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1126924	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1110058	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	445546	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	356161	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
35) Dibromofluoromethane	7.59	113	356236	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
50) Toluene-d8	10.09	98	1398738	52.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.62%	
62) 4-Bromofluorobenzene	12.40	95	491274	55.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.18%	
Target Compounds						
16) Acetone	3.82	43	23792	13.643	ug/l	97
20) Methylene Chloride	4.55	84	254834	34.832	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021919\
Data File : VN053862.D
Acq On : 19 Feb 2019 12:01
Operator : JC/SP
Sample : K1525-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
M-14-D

Quant Time: Feb 20 05:48:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

