

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021919\
 Data File : VN053864.D
 Acq On : 19 Feb 2019 12:55
 Operator : JC/SP
 Sample : K1518-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-MH-C05

Quant Time: Feb 20 05:49:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	659204	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1146488	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1119336	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	456686	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	364342	55.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.18%	
35) Dibromofluoromethane	7.59	113	360473	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
50) Toluene-d8	10.09	98	1411284	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
62) 4-Bromofluorobenzene	12.40	95	490498	54.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.14%	
Target Compounds						
16) Acetone	3.82	43	17512	10.324	ug/l	100
20) Methylene Chloride	4.56	84	27535	3.869	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021919\
Data File : VN053864.D
Acq On : 19 Feb 2019 12:55
Operator : JC/SP
Sample : K1518-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-MH-C05

Quant Time: Feb 20 05:49:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

