

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN021919\
 Data File : VN053866.D
 Acq On : 19 Feb 2019 13:48
 Operator : JC/SP
 Sample : K1518-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CABLE-BRIDGE-FOUNDATION-D-7

Quant Time: Feb 20 05:49:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	611751	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1092256	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	996705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	337980	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	327527	53.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.74%	
35) Dibromofluoromethane	7.59	113	331729	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
50) Toluene-d8	10.09	98	1267308	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
62) 4-Bromofluorobenzene	12.40	95	386832	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
Target Compounds						
16) Acetone	3.82	43	15106	9.596	ug/l	98
20) Methylene Chloride	4.55	84	24882	3.768	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN021919\
Data File : VN053866.D
Acq On : 19 Feb 2019 13:48
Operator : JC/SP
Sample : K1518-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-7

Quant Time: Feb 20 05:49:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

